



Eaton Analytical

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)



Laboratory Report

for

Horsham Water & Sewer Authority
617 Horsham Road
Horsham, PA 19044
Attention: Tina M. O'Rourke
Fax: (215) 672-8065

Date of Issue
07/09/2014


EUROFINS EATON
ANALYTICAL

TDF: Thomas.D.French
Project Manager

Report: 487430
Project: UCMR3
Group: PA1460033/Horsham
W&SA GW

* Accredited in accordance with TNI 2009 and ISO/IEC 17025:2005.

* Laboratory certifies that the test results meet all **TNI 2009 and ISO/IEC 17025:2005** requirements unless noted under the individual analysis.

* Following the cover page are State Certification List, ISO 17025 Accredited Method List, Acknowledgement of Samples Received, Comments, Hits Report, Data Report, QC Summary, QC Report and Regulatory Forms, as applicable.

* Test results relate only to the sample(s) tested.

STATE CERTIFICATION LIST

State	Certification Number	State	Certification Number
Alabama	41060	Mississippi	Certified
Alaska	CA00006	Montana	Cert 0035
Arizona	AZ0778	Nebraska	Certified
Arkansas	Certified	Nevada	CA00006-2014-1
California-Monrovia-ELAP	2813	New Hampshire *	2959
California-Colton- ELAP	2812	New Jersey *	CA 008
California-Folsom- ELAP	2820	New Mexico	Certified
Colorado	Certified	New York *	11320
Connecticut	PH-0107	North Carolina	06701
Delaware	CA 006	North Dakota	R-009
Florida *	E871024	Oregon (Primary AB) *	ORELAP 4034
Georgia	947	Pennsylvania *	68-565
Guam	14-003r	Rhode Island	LAO00326
Hawaii	Certified	South Carolina	87016
Idaho	Certified	South Dakota	Certified
Illinois *	200033	Tennessee	TN02839
Indiana	C-CA-01	Texas *	T104704230-14-6
Kansas *	E-10268	Utah *	CA000062014-7
Kentucky	90107	Vermont	VT0114
Louisiana *	LA140009	Virginia *	460260
Maine	CA0006	Washington	C838
Maryland	224	West Virginia	9943 C
Commonwealth of Northern Marianas Is.	MP0004	Wisconsin	998316660
Massachusetts	M-CA006	Wyoming	8TMS-L
Michigan	9906	EPA Region 5	Certified
Los Angeles County Sanitation Districts	10264		

* NELAP/TNI Recognized Accreditation Bodies

ISO 17025 Accredited Method List

The tests listed below are accredited and meet the requirements of ISO 17025 as verified by the ANSI-ASQ National Accreditation Board/ACLASS.
Refer to Certificate and scope of accreditation (AT 1807) found at: <http://www.eatonanalytical.com>

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Drinking Water	Food & Beverage	Waste Water
1,4-Dioxane	EPA 522	x	x	
2,3,7,8-TCDD	Modified EPA 1613B	x	x	
Acrylamide	In House Method	x	x	
Alkalinity	SM 2320B	x	x	x
Ammonia	EPA 350.1		x	x
Ammonia	SM 4500-NH3 H (18th)		x	x
Anions and DBPs by IC	EPA 300.0	x	x	x
Anions and DBPs by IC	EPA 300.1	x	x	
Asbestos	EPA 100.2	x		
Bicarbonate Alkalinity as HCO3	SM 2330B	x	x	x
BOD / CBOD	SM 5210B		x	x
Bromate	In House Method	x	x	
Carbamates	EPA 531.2	x	x	
Carbonate as CO3	SM 2330B	x	x	x
Carbonyls	EPA 556	x	x	
COD	EPA 410.4 / SM 5220D			x
Chloramines	SM 4500-CL G	x	x	x
Chlorinated Acids	EPA 515.4	x	x	
Chlorinated Acids	EPA 555	x	x	
Chlorine Dioxide	SM 4500-CLO2 D	x	x	
Chlorine -Total/Free/Combined Residual	SM 4500-Cl G	x	x	x
Conductivity	EPA 120.1			x
Conductivity	SM 2510B	x	x	x
Corrosivity (Langelier Index)	SM 2330B	x	x	
Cyanide, Amenable	SM 4500-CN G	x		x
Cyanide, Free	SM 4500CN F	x	x	x
Cyanide, Total	EPA 335.4	x	x	x
Cyanogen Chloride (screen)	In House Method	x	x	
Diquat and Paraquat	EPA 549.2	x	x	
DBP/HAA	SM 6251B	x	x	
Dissolved Oxygen	SM 4500-O G		x	x
E. Coli	(MTF/EC+MUG)	x		
E. Coli	CFR 141.21(f)(6)(i)		x	x
E. Coli	SM 9223			x
E. Coli (Enumeration)	SM 9221B.1/ SM 9221F	x	x	
E. Coli (Enumeration)	SM 9223B	x	x	
EDB/DCBP	EPA 504.1	x		
EDB/BCP and DBP	EPA 551.1	x	x	
EDTA and NTA	In House Method	x	x	
Endothall	EPA 548.1	x	x	
Enterococci	SM 9230B	x		x
Fecal Coliform	SM 9221 E (MTF/EC)	x		
Fecal Coliform	SM 9221C, E (MTF/EC)			x
Fecal Coliform (Enumeration)	SM 9221E (MTF/EC)	x	x	
Fecal Coliform with Chlorine Present	SM 9221E			x
Fecal Streptococci	SM 9230B	x		x
Fluoride	SM 4500-F C	x	x	x
Glyphosate	EPA 547	x	x	
Gross Alpha/Beta	EPA 900.0	x	x	x
HAAs/ Dalapon	EPA 552.3	x	x	
Hardness	SM 2340B	x	x	x
Heterotrophic Bacteria	In House Method	x	x	
Heterotrophic Bacteria	SM 9215 B	x	x	
Hexavalent Chromium	EPA 218.6	x	x	x
Hexavalent Chromium	EPA 218.7	x	x	
Hexavalent Chromium	SM 3500-Cr B or C (20th)			x

SPECIFIC TESTS	METHOD OR TECHNIQUE USED	Drinking Water	Food & Beverage	Waste Water
Hormones	EPA 539	x	x	
Hydroxide as OH Calc.	SM 2330B	x	x	
Kjeldahl Nitrogen	EPA 351.2			x
Mercury	EPA 245.1	x	x	x
Metals	EPA 200.7 / 200.8	x	x	x
Microcystin LR	ELISA	x	x	
NDMA	EPA 521	x	x	
Nitrate/Nitrite Nitrogen	EPA 353.2	x	x	x
OCL, Pesticides/PCB	EPA 505	x	x	
Ortho Phosphate	EPA 365.1	x	x	
Ortho Phosphate and Total Phosphorous	EPA 365.1/SM 4500-P E			x
Ortho Phosphorous	SM 4500P E	x	x	
Oxyhalides Disinfection Byproducts	EPA 317.0	x	x	
Perchlorate	EPA 331.0	x	x	
Perchlorate	EPA 314.0	x	x	
Perfluorinated Alkyl Acids	EPA 537	x	x	
pH	EPA 150.1	x		
pH	SM 4500-H+B	x	x	x
Phenylurea Pesticides/Herbicides	In House Method	x	x	
Pseudomonas	IDEXX Pseudalert	x	x	
Radium-226	RA-226 GA	x	x	
Radium-228	RA-228 GA	x	x	
Radon-222	SM 7500RN	x	x	
Residue, Filterable	SM 2540C	x	x	x
Residue, Non-filterable	SM 2540D			x
Residue, Total	SM 2540B		x	x
Residue, Volatile	EPA 160.4			x
Semi-VOC	EPA 525.2	x	x	
Semi-VOC	EPA 625	x	x	x
Silica	SM 4500-Si D	x	x	x
Silica	SM 4500-SiO2 C	x		x
Sulfide	SM 4500-S ⁻ D			x
Sulfite	SM 4500-SO ³ B	x	x	x
Surfactants	SM 5540C	x	x	x
Taste and Odor Analytes	SM 6040E	x	x	
Total Coliform	SM 9221 A, B	x	x	
Total Coliform (Enumeration)	SM 9221 A, B, C	x	x	
Total Coliform / E. coli	Colisure	x	x	
Total Coliform	SM 9221B			x
Total Coliform with Chlorine Present	SM 9221B			x
Total Coliform / E.coli	SM 9223	x	x	
TOC	SM 5310C		x	x
TOC/DOC	SM 5310C	x	x	
TOX	SM 5320B			x
Total Phenols	EPA 420.1			x
Total Phenols	EPA 420.4	x	x	x
Total Phosphorous	SM 4500 P F			x
Turbidity	EPA 180.1	x	x	x
Turbidity	SM 2130B	x		x
Uranium by ICP/MS	EPA 200.8	x	x	
UV 254	SM 5910B	x		
VOC	EPA 524.2/EPA 524.3	x	x	
VOC	EPA 624	x	x	x
VOC	EPA SW 846 8260	x	x	
VOC	In House Method	x	x	
Yeast and Mold	SM 9610	x	x	

750 Royal Oaks Dr., Ste 100, Monrovia, CA 91016 Tel (626) 386-1100 Fax (626) 386-1101 <http://www.EatonAnalytical.com>

Version 002. Issued: 06/03/2014

Acknowledgement of Samples Received

Addr: **Horsham Water & Sewer Authority**
 617 Horsham Road
 Horsham, PA 19044

Client ID: HORSHAM-PA
 Folder #: 487430
 Project: UCMR3
 Sample Group: PA1460033/Horsham W&SA GW

Attn: Tina M. O'Rourke
 Phone: (215) 672-8011

Project Manager: Thomas.D.French
 Phone: (480) 778-1558

The following samples were received from you on **June 26, 2014**. They have been scheduled for the tests listed below each sample. If this information is incorrect, please contact your service representative. Thank you for using Eurofins Eaton Analytical.


Sample #	Sample ID	Sample Date
<u>201406260436</u>	00112-112-Well # 20	06/24/2014 1100
	Sample Type: EP Sample Event: SE1 Facility ID: 00112 Sample Point ID: 112 PWSID: PA1460033 Static ID: EP	
	@UCMR3 522 C	@UCMR3 537
<u>201406260437</u>	FB::00112-112-Well # 20	06/24/2014 1100
	Static ID: FB	
	@UCMR3 537 FB	

Test Description

@UCMR3 522 C -- UCMR3 1,4-Dioxane by EPA 522

@UCMR3 537 -- UCMR3 537

@UCMR3 537 FB -- UCMR3 537



Eaton Analytical

750 Royal Oaks Drive, Suite 100, Monrovia, CA 91016-3629
Phone: 626 386 1100/ 800 566 LABS (800 566 5227)
Fax: 626 386 1101
Website: www.EatonAnalytical.com

UCMR3 CHAIN OF CUSTODY RECORD

487430

EUROFINS EATON ANALYTICAL USE ONLY:

Folder No:

LOGIN COMMENTS:

SAMPLES CHECKED AGAINST COC BY: AS

SAMPLES LOGGED IN BY: AS

SAMPLE TEMP RECEIVED AT:

Monrovia °C (Compliance: ≤10°C for the first 48 hours or ≤6°C after 48 hours)
4.8 °C (Compliance: ≤10°C for the first 48 hours or ≤6°C after 48 hours)

CONDITION OF BLUE ICE: Frozen X Thawed _____ No Ice _____

Pick-Up / Waik-In / FedEx / UPS / DHL / Area Fast / Top Line / Other: _____

PWSID: PA1460033

Example: (CA1234567)

TO BE COMPLETED BY SAMPLER:

COMPANY/AGENCY NAME: Aqua Pennsylvania, Inc.		PROJECT CODE: UCMR3-PA	COMPLIANCE SAMPLES - Requires upload to EPA database		Regulation: UCMR3
EEA CLIENT CODE: PHILLYSUB	COC ID:	SAMPLE GROUP: PA1460033/ Horsham Water & Sewer Authority	- Requires upload to EPA database		DTE: we MUST have PWSID#, Facility ID, Sample Point ID, and Sample Type to be able to upload data to EPA Database
TAT requested: rush by adv notice only		STD: <u>x</u> 1 wk _____ 3 day _____ 2 day _____ 1 day _____	SEE ATTACHED BOTTLE ORDER FOR ANALYSES		(check for yes), <u>OR</u>

SAMPLE DATE	SAMPLE TIME	FACILITY ID (per EPA Requirement) - 5 characters Max	UNIQUE FIELD SAMPLE ID (per EPA Requirement) - 20 characters max	WATER SOURCE TYPE #	SAMPLE EVENT #	SAMPLE POINT TYPE ID #	DISINFECTANT TYPE #	UCMR3 200.8	UCMR3 200.8 - FB	UCMR3 218.7	UCMR3 Chlorate	UCMR3 524.3	UCMR3 524.3 - TB	UCMR3 522	UCMR3 537	UCMR3 537 - FB	UCMR3 539	UCMR3 539 - FB	SAMPLER COMMENTS
06/24/14	11:00	00112-112	Well #20	GW	SE1	EP								X	X				

(1) Water Source Type: SW: Surface Water GW: Ground Water GU: Ground Water under the direct influence of SW MX: Any Combination of previous three water types
 (2) Sample Event Code: SE1 (first) SE2 (second) SE3 (third) SE4 (fourth)
 (3) Sampling Point Type ID: EP: Entry Point to the distribution system MR: Distribution System sample at maximum residence time

(4) Disinfectant Type:
 CLGA: Gaseous Chlorine CLOF: Offsite Generated Hypochlorite (stored as liquid form) CLON: Onsite Generated Hypochlorite (no storage) CAGC: Chloramine (formed from gaseous chlorine) CAOF: Chloramine (formed from onsite hypochlorite) CAON: Chloramine (formed from onsite hypochlorite) CLDO: Chlorine Dioxide OZON: Ozone ULVL: Ultraviolet Light OTHD: All Other Types of Disinfectant NODU: No Disinfectant Used

* Field Blank (FB) or Trip Blank (TB) are analyzed only when associated samples have positive results (>MRL)

SAMPLED BY:	SIGNATURE	PRINT NAME	COMPANY	DATE	TIME
RELINQUISHED BY:		B. Pachik	Aqua PA	06/24/14	11:00
RECEIVED BY:					
RELINQUISHED BY:		A. Nguyen	Aqua PA	06/25/14	16:00
RECEIVED BY:		VICTOR PLASENCIA	EEA	06/26/14	1146

From: (610) 645-1176
Michael Senft
Aqua
762 West Lancaster Avenue

Bryn Mawr, PA 19010

Origin ID: WAYA



J111011402070326

Ship Date: 25JUN14
ActWgt: 15.0 LB
CAD: 100061330/NET3490
Dims: 24 X 13 X 14 IN

Delivery Address Bar Code



SHIP TO: (626) 386-1100
LOGIN
MWH Americas,INC,
750 Royal Oaks Drive
Suite 100
Monrovia, CA 91016

BILL RECIPIENT

Ref # UCMR3
Invoice # 15-3031
PO #
Dept # 15-3031

4 of 4

THU - 26 JUN AA
STANDARD OVERNIGHT

MPS# 7704 1278 2216

0263

Mstr# 7704 1278 2010



0201

91016

CA-US

BUR

NC WHPA



522G5/B04/F220

After printing this label:

1. Use the 'Print' button on this page to print your label to your laser or inkjet printer.
2. Fold the printed page along the horizontal line.
3. Place label in shipping pouch and affix it to your shipment so that the barcode portion of the label can be read and scanned.

Warning: Use only the printed original label for shipping. Using a photocopy of this label for shipping purposes is fraudulent and could result in additional billing charges, along with the cancellation of your FedEx account number.

Use of this system constitutes your agreement to the service conditions in the current FedEx Service Guide, available on fedex.com. FedEx will not be responsible for any claim in excess of \$100 per package, whether the result of loss, damage, delay, non-delivery, misdelivery, or misinformation, unless you declare a higher value, pay an additional charge, document your actual loss and file a timely claim. Limitations found in the current FedEx Service Guide apply. Your right to recover from FedEx for any loss, including intrinsic value of the package, loss of sales, income interest, profit, attorney's fees, costs, and other forms of damage whether direct, incidental, consequential, or special is limited to the greater of \$100 or the authorized declared value. Recovery cannot exceed actual documented loss. Maximum for items of extraordinary value is \$1,000, e.g. jewelry, precious metals, negotiable instruments and other items listed in our ServiceGuide. Written claims must be filed within strict time limits, see current FedEx Service Guide.



Eaton Analytical

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Comments
Report: 487430

Horsham Water & Sewer Authority
Tina M. O'Rourke
617 Horsham Road
Horsham, PA 19044

The Comments Report may be blank if there are no comments for this report.



Eaton Analytical

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

UCMR Field Blank
Report: 487430

Horsham Water & Sewer Authority

Tina M. O'Rourke
617 Horsham Road
Horsham, PA 19044

Samples Received on:
06/26/2014

The results section will be blank if there are no exceedances of UCMR3 Field Blank criteria. Field Blank Evaluation is required for positive detection in the associated sample for Metals by 200.8, VOCs by 524.3, PFCs by 537, and Hormones by 539 (SS monitoring only). A detection on this report indicates need for re-sample for the associated site and test. Reference: UCMR3 Laboratory Approval Requirements and Information Document V2, May 2012 section 8.1 Field Blanks.

UCMR3 Field Blanks are not required to be analyzed, if the target analytes are not detected in the associated samples. In that event, the Field Blank data are not available (NA) for reporting.

Results Section

Analyzed	Analyte	Sample ID	Result	Units	UCMR Limit
----------	---------	-----------	--------	-------	------------

SUMMARY OF POSITIVE DATA ONLY

<u>Method</u>	<u>Method description</u>	<u>Positive Data Limit = UCMR Limit</u>
@UCMR3 200.8 FB	Metals	Any detection Greater than 1/3
@UCMR3 524.3 TB	Volatiles	MRL
@UCMR3 537 FB	Perfluorinated	Any detection Greater than 1/3
@UCMR3 539 FB	Hormones	MRL



Eaton Analytical

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Hits
Report: 487430

Horsham Water & Sewer Authority
Tina M. O'Rourke
617 Horsham Road
Horsham, PA 19044

Samples Received on:
06/26/2014

Analyzed	Analyte	Sample ID	Result	Federal MCL	Units	MRL
----------	---------	-----------	--------	-------------	-------	-----

SUMMARY OF POSITIVE DATA ONLY



Eaton Analytical

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Laboratory Data
Report: 487430

Horsham Water & Sewer Authority
Tina M. O'Rourke
617 Horsham Road
Horsham, PA 19044

Samples Received on:
06/26/2014

Prepared	Analyzed	QC Ref #	Method	Analyte	Result	Units	MRL	Dilution
00112-112-Well # 20 (201406260436)						Sampled on 06/24/2014 1100		
Sample Type: EP								
Sample Event: SE1								
Facility ID: 00112								
Sample Point ID: 112								
PWSID: PA1460033								
Static ID: EP								
EPA 522 - UCMR3 1,4-Dioxane by EPA 522								
6/30/2014	07/03/2014	20:14	779457	(EPA 522)	1,4-Dioxane	ND	ug/L	0.07 1
6/30/2014	07/03/2014	20:14	779457	(EPA 522)	Dioxane-d8	92	%	1
6/30/2014	07/03/2014	20:14	779457	(EPA 522)	THF-d8	75	%	1
EPA 537 - UCMR3 537								
07/01/2014	19:20	778753	(EPA 537)	Perfluoro octanesulfonic acid - PFOS	ND	ug/L	0.04	1
07/01/2014	19:20	778753	(EPA 537)	Perfluoro-1-butanesulfonic acid -PFBS	ND	ug/L	0.09	1
07/01/2014	19:20	778753	(EPA 537)	Perfluoro-1-hexanesulfonic acid - PFHxS	ND	ug/L	0.03	1
07/01/2014	19:20	778753	(EPA 537)	Perfluoroheptanoic acid - PFHpA	ND	ug/L	0.01	1
07/01/2014	19:20	778753	(EPA 537)	Perfluoro-n-nonanoic acid -PFNA	ND	ug/L	0.02	1
07/01/2014	19:20	778753	(EPA 537)	Perfluorooctanoic acid - PFOA	ND	ug/L	0.02	1
07/01/2014	19:20	778753	(EPA 537)	13C-PFDA - Surr#2	82	%		1
07/01/2014	19:20	778753	(EPA 537)	13C-PFHxA - Surr#1	84	%		1
07/01/2014	19:20	778753	(EPA 537)	13C-PFOA- IS#1	122	%		1
07/01/2014	19:20	778753	(EPA 537)	13C-PFOS- IS#2	119	%		1

Rounding on totals after summation.
(c) - indicates calculated results



Eaton Analytical

Laboratory
QC Summary: 487430

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Horsham Water & Sewer Authority

QC Ref # 778753 - UCMR3 537

201406260436 00112-112-Well # 20

Analysis Date: 07/01/2014

Analyzed by: 1CL

QC Ref # 779457 - UCMR3 1,4-Dioxane by EPA 522

201406260436 00112-112-Well # 20

Analysis Date: 07/03/2014

Analyzed by: CRW

750 Royal Oaks Drive, Suite 100
Monrovia, California 91016-3629
Tel: (626) 386-1100
Fax: (626) 386-1101
1 800 566 LABS (1 800 566 5227)

Horsham Water & Sewer Authority

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
QC Ref# 778753 - UCMR3 537 by EPA 537						Analysis Date: 07/01/2014			
CCCH	13C-PFDA - Surr#2 (S)			101	%	101	(70-130)		
CCCL	13C-PFDA - Surr#2 (S)			99.3	%	99	(70-130)		
CCCM	13C-PFDA - Surr#2 (S)			97.8	%	98	(70-130)		
MBLK_HI	13C-PFDA - Surr#2 (S)			86.4	%	86	(70-130)		
MRLHI	13C-PFDA - Surr#2 (S)			86.6	%	87	(70-130)		
MS2_201406260469	13C-PFDA - Surr#2 (S)			89.8	%	90	(70-130)		
MSD2_201406260469	13C-PFDA - Surr#2 (S)			85.7	%	86	(70-130)		
QCS	13C-PFDA - Surr#2 (S)			102	%	102	(70-130)		
CCCH	13C-PFHxA - Surr#1 (S)			101	%	101	(70-130)		
CCCL	13C-PFHxA - Surr#1 (S)			100	%	100	(70-130)		
CCCM	13C-PFHxA - Surr#1 (S)			101	%	101	(70-130)		
MBLK_HI	13C-PFHxA - Surr#1 (S)			90.6	%	91	(70-130)		
MRLHI	13C-PFHxA - Surr#1 (S)			91.7	%	92	(70-130)		
MS2_201406260469	13C-PFHxA - Surr#1 (S)			81.9	%	82	(70-130)		
MSD2_201406260469	13C-PFHxA - Surr#1 (S)			80.6	%	81	(70-130)		
QCS	13C-PFHxA - Surr#1 (S)			101	%	101	(70-130)		
CCCH	13C-PFOA- IS#1 (I)			101	%	101	(50-150)		
CCCL	13C-PFOA- IS#1 (I)			101	%	101	(50-150)		
CCCM	13C-PFOA- IS#1 (I)			98.4	%	98	(50-150)		
MBLK_HI	13C-PFOA- IS#1 (I)			122	%	122	(50-150)		
MRLHI	13C-PFOA- IS#1 (I)			119	%	119	(50-150)		
MS2_201406260469	13C-PFOA- IS#1 (I)			114	%	114	(50-150)		
MSD2_201406260469	13C-PFOA- IS#1 (I)			116	%	116	(50-150)		
QCS	13C-PFOA- IS#1 (I)			96.5	%	97	(50-150)		
CCCH	13C-PFOS- IS#2 (I)			103	%	103	(50-150)		
CCCL	13C-PFOS- IS#2 (I)			100	%	100	(50-150)		
CCCM	13C-PFOS- IS#2 (I)			101	%	101	(50-150)		
MBLK_HI	13C-PFOS- IS#2 (I)			123	%	123	(50-150)		
MRLHI	13C-PFOS- IS#2 (I)			118	%	119	(50-150)		
MS2_201406260469	13C-PFOS- IS#2 (I)			117	%	117	(50-150)		
MSD2_201406260469	13C-PFOS- IS#2 (I)			117	%	117	(50-150)		
QCS	13C-PFOS- IS#2 (I)			99.5	%	99	(50-150)		
CCCH	Perfluoro octanesulfonic acid - PFOS		0.13	0.121	ug/L	95	(70-130)		
CCCL	Perfluoro octanesulfonic acid - PFOS		0.032	0.0310	ug/L	97	(50-150)		
CCCM	Perfluoro octanesulfonic acid - PFOS		0.064	0.0613	ug/L	96	(70-130)		
MBLK_HI	Perfluoro octanesulfonic acid - PFOS	ND		<0.01333	ug/L				

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

750 Royal Oaks Drive, Suite 100
 Monrovia, California 91016-3629
 Tel: (626) 386-1100
 Fax: (626) 386-1101
 1 800 566 LABS (1 800 566 5227)

Laboratory QC
 Report: 487430

Horsham Water & Sewer Authority

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
MRLHI	Perfluoro octanesulfonic acid - PFOS		0.04	0.0340	ug/L	85	(50-150)		
MS2_201406260469	Perfluoro octanesulfonic acid - PFOS	ND	0.08	0.0798	ug/L	89	(70-130)		
MSD2_201406260469	Perfluoro octanesulfonic acid - PFOS	ND	0.08	0.0794	ug/L	89	(70-130)	30	0.50
QCS	Perfluoro octanesulfonic acid - PFOS		0.048	0.0472	ug/L	99	(70-130)		
CCCH	Perfluoro-1-butanefulfonic acid -PFBS		0.29	0.282	ug/L	97	(70-130)		
CCCL	Perfluoro-1-butanefulfonic acid -PFBS		0.078	0.0771	ug/L	99	(50-150)		
CCCM	Perfluoro-1-butanefulfonic acid -PFBS		0.15	0.149	ug/L	102	(70-130)		
MBLK_HI	Perfluoro-1-butanefulfonic acid -PFBS	ND		<0.03033	ug/L				
MRLHI	Perfluoro-1-butanefulfonic acid -PFBS		0.09	0.0851	ug/L	95	(50-150)		
MS2_201406260469	Perfluoro-1-butanefulfonic acid -PFBS	ND	0.18	0.177	ug/L	96	(70-130)		
MSD2_201406260469	Perfluoro-1-butanefulfonic acid -PFBS	ND	0.18	0.176	ug/L	96	(70-130)	30	0.57
QCS	Perfluoro-1-butanefulfonic acid -PFBS		0.044	0.0496	ug/L	112	(70-130)		
CCCH	Perfluoro-1-hexanesulfonic acid - PFHxS		0.096	0.0899	ug/L	94	(70-130)		
CCCL	Perfluoro-1-hexanesulfonic acid - PFHxS		0.024	0.0238	ug/L	99	(50-150)		
CCCM	Perfluoro-1-hexanesulfonic acid - PFHxS		0.048	0.0447	ug/L	93	(70-130)		
MBLK_HI	Perfluoro-1-hexanesulfonic acid - PFHxS	ND		<0.0100	ug/L				
MRLHI	Perfluoro-1-hexanesulfonic acid - PFHxS		0.03	0.0262	ug/L	87	(50-150)		
MS2_201406260469	Perfluoro-1-hexanesulfonic acid - PFHxS	ND	0.06	0.0587	ug/L	92	(70-130)		
MSD2_201406260469	Perfluoro-1-hexanesulfonic acid - PFHxS	ND	0.06	0.0584	ug/L	91	(70-130)	30	0.51
QCS	Perfluoro-1-hexanesulfonic acid - PFHxS		0.048	0.0448	ug/L	94	(70-130)		
CCCH	Perfluoroheptanoic acid - PFHpA		0.032	0.0317	ug/L	99	(70-130)		
CCCL	Perfluoroheptanoic acid - PFHpA		0.008	0.00826	ug/L	103	(50-150)		
CCCM	Perfluoroheptanoic acid - PFHpA		0.016	0.0158	ug/L	99	(70-130)		
MBLK_HI	Perfluoroheptanoic acid - PFHpA	ND		<0.00333	ug/L				
MRLHI	Perfluoroheptanoic acid - PFHpA		0.01	0.00881	ug/L	88	(50-150)		
MS2_201406260469	Perfluoroheptanoic acid - PFHpA	ND	0.02	0.0214	ug/L	90	(70-130)		
MSD2_201406260469	Perfluoroheptanoic acid - PFHpA	ND	0.02	0.0212	ug/L	89	(70-130)	30	0.94
QCS	Perfluoroheptanoic acid - PFHpA		0.05	0.0543	ug/L	109	(70-130)		
CCCH	Perfluoro-n-nonanoic acid -PFNA		0.064	0.0639	ug/L	100	(70-130)		
CCCL	Perfluoro-n-nonanoic acid -PFNA		0.016	0.0171	ug/L	107	(50-150)		
CCCM	Perfluoro-n-nonanoic acid -PFNA		0.032	0.0337	ug/L	105	(70-130)		
MBLK_HI	Perfluoro-n-nonanoic acid -PFNA	ND		<0.00666	ug/L				
MRLHI	Perfluoro-n-nonanoic acid -PFNA		0.02	0.0185	ug/L	92	(50-150)		
MS2_201406260469	Perfluoro-n-nonanoic acid -PFNA	ND	0.04	0.0388	ug/L	93	(70-130)		
MSD2_201406260469	Perfluoro-n-nonanoic acid -PFNA	ND	0.04	0.0384	ug/L	92	(70-130)	30	1.0
QCS	Perfluoro-n-nonanoic acid -PFNA		0.05	0.0490	ug/L	98	(70-130)		
CCCH	Perfluorooctanoic acid - PFOA		0.064	0.0631	ug/L	99	(70-130)		
CCCL	Perfluorooctanoic acid - PFOA		0.016	0.0163	ug/L	102	(50-150)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.

750 Royal Oaks Drive, Suite 100
 Monrovia, California 91016-3629
 Tel: (626) 386-1100
 Fax: (626) 386-1101
 1 800 566 LABS (1 800 566 5227)

Laboratory QC
 Report: 487430

Horsham Water & Sewer Authority

QC Type	Analyte	Native	Spiked	Recovered	Units	Yield (%)	Limits (%)	RPDLimit (%)	RPD%
CCCM	Perfluorooctanoic acid - PFOA		0.032	0.0326	ug/L	102	(70-130)		
MBLK_HI	Perfluorooctanoic acid - PFOA	ND		<0.00666	ug/L				
MRLHI	Perfluorooctanoic acid - PFOA		0.02	0.0184	ug/L	92	(50-150)		
MS2_201406260469	Perfluorooctanoic acid - PFOA	ND	0.04	0.0444	ug/L	95	(70-130)		
MSD2_201406260469	Perfluorooctanoic acid - PFOA	ND	0.04	0.0434	ug/L	93	(70-130)	30	2.3
QCS	Perfluorooctanoic acid - PFOA		0.05	0.0511	ug/L	102	(70-130)		
QC Ref# 779457 - UCMR3 1,4-Dioxane by EPA 522 by EPA 522						Analysis Date: 07/03/2014			
CCCH	1,4-Dioxane		50	53.6	ug/L	107	(70-130)		
CCCL	1,4-Dioxane		0.07	0.0670	ug/L	96	(50-150)		
CCCM	1,4-Dioxane		20	20.9	ug/L	105	(70-130)		
LCS1	1,4-Dioxane		20	16.8	ug/L	84	(70-130)		
MBLK	1,4-Dioxane			<0.023	ug/L				
MRL_CHK	1,4-Dioxane		0.07	0.0680	ug/L	97	(50-150)		
MS_201406250115	1,4-Dioxane	ND	20	16.9	ug/L	84	(70-130)		
MSD_201406250115	1,4-Dioxane	ND	20	16.7	ug/L	84	(70-130)	20	1.2
CCCH	Dioxane-d8			112	%	113	(70-130)		
CCCL	Dioxane-d8			106	%	106	(70-130)		
CCCM	Dioxane-d8			115	%	115	(70-130)		
LCS1	Dioxane-d8			90.3	%	90	(70-130)		
MBLK	Dioxane-d8			88.6	%				
MRL_CHK	Dioxane-d8			91.2	%	91	(70-130)		
MS_201406250115	Dioxane-d8	90		89.9	%	90	(70-130)		
MSD_201406250115	Dioxane-d8	90		89.8	%	90	(70-130)		
CCCH	THF-d8			75.9	%	76	(50-150)		
CCCL	THF-d8			77.2	%	77	(50-150)		
CCCM	THF-d8			71.0	%	71	(50-150)		
LCS1	THF-d8			68.9	%	69	(50-150)		
MBLK	THF-d8			69.9	%				
MRL_CHK	THF-d8			69.5	%	70	(50-150)		
MS_201406250115	THF-d8	71		65.7	%	66	(50-150)		
MSD_201406250115	THF-d8	71		75.5	%	76	(50-150)		

Spike recovery is already corrected for native results.

Spikes which exceed Limits and Method Blanks with positive results are highlighted by Underlining.

Criteria for MS and Dup are advisory only, batch control is based on LCS. Criteria for duplicates are advisory only, unless otherwise specified in the method.

RPD not calculated for LCS2 when different a concentration than LCS1 is used.

RPD not calculated for Duplicates when the result is not five times the MRL (Minimum Reporting Level).

(S) - Indicates surrogate compound.

(I) - Indicates internal standard compound.